

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

JAN 03 2012

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N	130526
Name		Joseph V. Stoenmaker	ID# 4188 Date 12-28-11
A	Agency	SDP1 (SAND POINT P.D.)	907-383-3700
Instrument Location		SAND POINT, AK	
Alco S/N		3B	Target Value .080 High Pressure 900
Alco Test Value Average		.076	.077
		1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature		Joseph V. Stoenmaker	
		(OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 2nd day of Feb, 2012 uk

Nikki Roth (Notary Seal Stamp)  
Daniel L. Nicholas  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

13 0526

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130526

DECEMBER 28, 2011

OPERATOR'S NAME:

SHOENAKER JOSEPH V

OPERATOR'S NUMBER: 4108

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

P

O.L. #: P

DEPT/AGENCY: SDP1

CASE/REPORT: P

TEST TYPE: V

ALCO TARGET VALUE: .000

ALCO S/N: 130526

NOB 28/12 3B  
BREATH ANALYSIS ---

.000 ADJUSTED FOR 29.65 in

ALCO TARGET .079 20:31

BLANK TEST .000 20:33

INTERNAL STANDARD VERIFIED 20:33

ALCO TV 29.65 in .076 20:33

BLANK TEST .000 20:34

SUBJECT SAMPLE .000 20:34

BLANK TEST .000 20:35

ALCO TV 29.65 in .077 20:35

BLANK TEST .000 20:36

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130526

DECEMBER 28, 2011

TIME 20:38

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 40c

BAROMETER: 29.65 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~